

Review Summary – CH370 - Exam 3

Mass Spectrometry: produces ions / uses electric and magnetic fields to measure the mass / charge ratio of the charged particles: Parts: ion source; analyzer; detector

Source: Electron impact (EI) / Chemical Ionization (CI) / Fast atom bombardment (FAB) / Field desorption (FD) / Electrospray ionization (ESI) / Laser desorption (LD)

Analyzer: ions separated according to mass. Quadrupole / Magnetic Sector / TOF

Detector which produces a signal from the separated ions.

Linked Systems: GC/MS; LC/MS; MS/MS

Time-of Flight (TOF) Mass Spectrometer: particles have same kinetic energy $KE = zV$ or $(Ze)Es$ where "Ze" is the charge, "E" the electric field, and "s" the length of the source region before particle is allowed to "drift" to the detector;

$$(m/Z) = 2eEs(t/D)^2, \text{ or } m = [2eEs(1/D)^2] Z t^2$$

$$\text{mass} = (\text{constant}) \times Z \times t^2$$

Source of "ions" - Applications with Biomacromolecules –

a) **Matrix-Assisted Laser Desorption-Ionization (MALDI) / TOF**


b) **Electrospray Ionization (ESI):** nondestructive / microdroplets

c) **Capillary Electrophoresis (CE) and ESI:** very small samples – femtomole (10^{-15})

Sequence Analysis Using Mass Spectrometry: MS/MS

Ligand Binding:

a) **Basic Equations:**



If $S_o \gg E_o$, then $[S] \sim [S_o]$ then $K_d [ES] = [E_o - ES][S_o]$

$$[ES] = E_o S_o / (K_d + S_o); \quad \text{define } \theta = [ES]/E_o = S_o / (K_d + S_o)$$

thus **$K_d = [S_o]$ when $\theta = 0.5$**

when above assumptions are not true \rightarrow solve quadratic equation (see notes)

$$\text{then } \theta = ((E_o + S_o + K_d) - \text{SQRT}((E_o + S_o + K_d)^2 - 4 E_o \times S_o)) / 2 E_o$$

b) Manipulations of Equation / **Plots**

1) Preferred – Non-linear regression analysis using statistical program

2) Common Plots:

Double reciprocal plot **$1/\theta = K_d/[S] + 1$** ; plot $1/\theta$ vs. $1/[S]$, slope = K_d

Scatchard Plot: $\theta = [S]/(K_d + [S])$ or $\theta K_d + \theta[S] = [S]$ or **$\theta = 1 - \theta K_d/[S]$**

plot θ vs. θ/S_o slope = $-K_d$

Multiple Binding Sites: define $v = n\theta$

$v = n - vK_d/[S]$ \rightarrow plot v vs. $v/[S]$; intercept = n and slope = $-K_d$

c) Measurements – how do experimentally determine $[S]$ or $[ES]$, etc:

1) **Equilibrium Dialysis** – Measure $[S]$ free (outside) and $[S]$ free + bound (inside)

2) Take advantage of difference in **fluorescence signal** between bound and unbound ligand species if possible. Such systems are more convenient to work with and are also readily adaptable to using with stopped flow kinetics to measure rate constants.

3) ITC – **Isothermal Calorimetry**

- a more rigorous alternative to van't Hoff analysis to obtain thermodynamic info.

- measure Δq 's directly during titration $\rightarrow n, \Delta H$ and K_{eq} from best fit to data.

$$K_{eq} \rightarrow \Delta G^\circ; \Delta H^\circ \text{ and } \Delta G^\circ \rightarrow \Delta S^\circ$$

4) SPR – **Surface Plasmon Resonance** – a quantum optical-electrical phenomena.

= Biacore (Biomolecular Interaction Analysis) - SPR can provide $k(\text{on})$ and $k(\text{off})$ data to give more than just K_{eq} values.

The detection principle relies on surface plasmon resonance (SPR), an electron charge density wave phenomenon that arises at the surface of a metallic film when light is reflected at the film under specific conditions. The resonance is a result of energy and momentum being transformed from incident photons into surface plasmons, and is sensitive to the refractive index of the medium on the opposite side of the film from the reflected light. The angle at which the light energy can be transferred (resonance coupled) to a metal (gold) surface depends on the refractive index of the material on the opposite side of the reflected light which in turn depends on the concentrations of receptor and bound ligand at the surface. By monitoring the SPR-angle as a function of time, the kinetic events at the surface can be displayed as a “**sensorgram**.”

- 5) Kinetics - Use rapid kinetics and a “mechanism” to model/fit **individual rate constants** of the chemical steps – more info that just k(on) and k(off) or Kd

X-ray: (refer to notes and graded homework for sample questions)

- Image Formation – Relationship of Resolution and Wavelength (role of amplitude and phase)
- Crystals and Crystal Growth / Crystal Lattices / Lattice Constants / Space Groups
- X-ray Sources – Sealed Tube / Rotation Anode vs. Synchrotron
Advantages of synchrotron radiation
- Diffraction - Bragg Equation - $n\lambda = 2d \sin(\theta)$ or just $\lambda = 2d_{hkl} \sin(\theta_{hkl})$
- Real Space Lattice/cell vs. Reciprocal Space Lattice/cell / Planes and indices
- Data Collection / Structure Factors: $F(hkl) = \text{SQRT}[I(hkl)]$
- Phase Problem: MIR / MR / MAD
(MIR) Multiple Isomorphous Replacement (Heavy Atom Method)
(MR) Molecular Replacement (Rotation and Translation of Model Structure)
(MAD) Multi-Wavelength Anomalous Dispersion Methods – 3 wavelengths over an absorption edge
- Refinement – simulated annealing using known bond distances, angles, etc. plus the agreement between calculated and observed structure factors for all reflections.

NMR: (Multidimensional NMR to determine structure via distance restraints)

- Spin:** Spin is a fundamental property of nature like electrical charge or mass (for a proton H; $I = 1/2$ and $m = \pm 1/2$). When a group of spins is placed in a magnetic field, each spin aligns in one of the two possible orientations ($\pm 1/2$). Magnetic field strength is measured in Tesla (T), vs. the earth's magnetic field in New York of approximately 5×10^{-5} T. A typical NMR spectrophometer has a magnetic field of 11.7 T. Nuclei with $I \neq 0$ will interact with an applied magnetic field, giving rise to quantized states where

$$E = \gamma h m H / 2\pi \text{ or } \Delta E = \gamma h H / 2\pi.$$

- NMR spectroscopy:** At room temperature, the number of spins in the lower energy level, N^+ , only slightly outnumbers the number in the upper level, N^- . **Boltzmann statistics** tells us that $N^-/N^+ = e^{-\Delta E/kT}$.

ΔE is the energy difference between the spin states; k is Boltzmann's constant, 1.3805×10^{-23} J/Kelvin; and T is the temperature in Kelvin.

The signal in NMR spectroscopy results from the difference between the energy absorbed by the spins which make a transition from the lower energy state to the higher energy state, and the energy emitted by the spins which simultaneously make a transition from the higher energy state to the lower energy state. The signal is thus proportional to the population

difference between the states. It is the resonance, or exchange of energy at a specific frequency between the spins and the spectrometer, which gives NMR its sensitivity.

$$\Delta E = \gamma h H / 2\pi = h \nu \quad \text{or transitions at } \nu = \gamma H / 2\pi$$

- c) **Chemical Shifts:** The total magnetic field felt by a nucleus is equal to the field applied plus any local field effects ($H_{\text{tot}} = H_{\text{applied}} + H_{\text{local}}$). Thus protons are first separated by general type ($-\text{CH}_3$ vs. $-\text{CH}_2-$ vs. aromatic vs. $-\text{N}-\text{H}$, etc.) by different chemical shifts (ppm).
- d) **Spin-spin coupling / Nuclear Overhauser Effect; NOEs::** Nuclei experiencing different environment or having different chemical shifts are nonequivalent. Nuclei which are close to one another exert an influence on each other's effective magnetic field. This effect shows up in the NMR spectrum when the nuclei are nonequivalent. If the distance between non-equivalent nuclei is less than or equal to three bond lengths, this effect is observable. This effect is called spin-spin coupling or J coupling. Nuclei can also show a through space coupling interaction. 2D spectra - irradiate at the frequency of one nuclei and look for an effect in another, nearby nucleus. Presence of **NOE peak indicates two nuclei that are within about 6Å** of each other. Need to be able to identify the peaks to use this information (“Assignment Problem” – see below).
- e) **Assignment Problem:** 3D spectra - can use ^{13}C and ^{15}N isotopically enriched proteins to spread out peaks to aid in making peak assignments. Make a list of assigned NOE peaks with labeled nuclei identified. This works for proteins up to about 250-300 amino acids.
- f) **Refinement by Simulated annealing:** Use molecular dynamics to minimize an “energy” function to arrive at a structure with lowest “energy” that gives good geometry for bond distances and bond angles, etc. plus good agreement with assigned list of NOE’s for close contacts. Molecules are allowed to sample more of conformational space by “heating” the sample in the computer by simulating higher temperatures and then slowly “cooled” in the computer simulation to be trapped in lowest minimum.
- g) **Presentation of NMR results:** NMR spectroscopists usually start with several starting models and “solve” or calculate an ensemble of final structural models. These are reported in two ways: 1) “**Blur-o-gram**” and 2) a plot of rmsd’s vs. amino acid residue number.

.....
Add For Final Exam:

Microscopy in Biochemistry and Medicine

TEM vs. SEM – principles behind images, resolution, type of results obtained

STM or AFM - principles behind images, resolution, type of results obtained

CAT / MRI / PET – principles, applications, advantages and disadvantages of each method

Imaging using ultrasound